

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135848

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.091	0.099	92	85-115	05/20/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/20/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/20/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	05/20/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	05/20/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135851

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	05/20/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	05/20/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	05/20/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135852

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	05/20/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	05/20/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	05/20/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	05/20/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135884

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
ASTM COD	mg/L	49.668	50	99	95-105	05/20/2025
Sample ID: CCV1						
ASTM COD	mg/L	48.673	50	97	95-105	05/22/2025
Sample ID: CCV2						
ASTM COD	mg/L	51.657	50	103	95-105	05/22/2025
Sample ID: CCV3						
ASTM COD	mg/L	50.662	50	101	95-105	05/22/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135885

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	1	1	100	90-110	05/22/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	1	1	100	90-110	05/22/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	1	1	100	90-110	05/22/2025
Sample ID: CCV3 ASTM Ammonia	mg/L	1	1	100	90-110	05/22/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135885

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------