

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135848

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	05/20/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	05/20/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/20/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	05/20/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	0.00097	0.0025	J	0.00096	0.005	05/20/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135884

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/20/2025
Sample ID: CCB1 ASTM COD	mg/L	1.925	5.0000	J	1.50	10	05/22/2025
Sample ID: CCB2 ASTM COD	mg/L	1.925	5.0000	J	1.50	10	05/22/2025
Sample ID: CCB3 ASTM COD	mg/L	1.925	5.0000	J	1.50	10	05/22/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135885

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	05/22/2025
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	05/22/2025
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	05/22/2025
Sample ID: CCB3 ASTM Ammonia	mg/L	0.032	0.0500	J	0.030	0.1	05/22/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135885

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------