

Preparation Blank Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB135845BL TS	%	< 2.5000	2.5000	U	1	5	05/20/2025
Sample ID: LB135846BL TVS	%	< 5.0000	5.0000	U	1	10	05/20/2025
Sample ID: LB135872BL ASTM Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	05/21/2025
Sample ID: LB135879BL ASTM TS	mg/L	1	2.5000	J	1	5	05/22/2025
Sample ID: LB135884BL ASTM COD	mg/L	< 5.0000	5.0000	U	1.5	10.0	05/22/2025
Sample ID: LB135900BL Oil and Grease	mg/Kg	< 12.5000	12.5000	U	5.8	25	05/23/2025
Sample ID: PB168061BL Reactive Cyanide	mg/Kg	< 0.0250	0.0250	U	0.0084	0.05	05/20/2025
Sample ID: PB168070BL Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	05/20/2025
Sample ID: PB168114BL ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.03	0.1	05/22/2025
Sample ID: PB168114TB ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.03	0.1	05/22/2025