

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135872

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135872BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	05/21/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135884

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135884BS							
ASTM COD	mg/L	50	49.7		99	1	90-110	05/22/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135900

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135900BS							
Oil and Grease	mg/Kg	100	94.8		95	1	80-120	05/23/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2078

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135885

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168114BS							
ASTM Ammonia	mg/L	1	0.99		99	1	90-110	05/22/2025