


LB135895

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5746e-004

C= 8.7953e-002

Rho= 0.9993994

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	-180	0.000	-180	0			
0.2	0.200	0.226	0.026	536	0.0 %	536				
2.5	2.500	2.286	-0.214	8537	0.0 %	8537				
5.0	5.000	5.148	0.148	19653	0.0 %	19653				
7.5	7.500	7.618	0.118	29246	0.0 %	29246				
10.0	10.000	9.881	-0.119	38037	0.0 %	38037				

0/1.1
13
-9
3
2
-1

LB135895 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2025 08:53:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	20	-180	-	0.0000	7470A	PPB	23 May 2025 08:53:36	S	Std
0.2 - 1	50-2	536	-	0.2000	7470A	PPB	23 May 2025 08:55:54	S	Std
2.5 - 1	52.5	8537	-	2.5000	7470A	PPB	23 May 2025 08:58:11	S	Std
5.0 - 1	55	19653	-	5.0000	7470A	PPB	23 May 2025 09:03:18	S	Std
7.5 - 1	57.5	29246	-	7.5000	7470A	PPB	23 May 2025 09:05:34	S	Std
10.0 - 1	59.5	38037	-	10.0000	7470A	PPB	23 May 2025 09:10:49	S	Std
ICV15 - 1	ICV15	13908	3.6687	-	7470A	PPB	23 May 2025 09:13:44	U	SMPL
ICB15 - 1	ICB15	-426	-0.0217	-	7470A	PPB	23 May 2025 09:16:00	U	SMPL
CCV55 - 1	CCV55	20374	5.3335	-	7470A	PPB	23 May 2025 09:18:18	U	SMPL
CCB55 - 1	CCB55	-565	-0.0575	-	7470A	PPB	23 May 2025 09:20:34	U	SMPL
CRA - 1	CRA	443	0.2020	-	7470A	PPB	23 May 2025 09:22:51	U	SMPL
HighStd - 1	HighStd	38127	9.9042	-	7470A	PPB	23 May 2025 09:25:06	U	SMPL
ChkStd - 1	ChkStd	27884	7.2670	-	7470A	PPB	23 May 2025 09:27:24	U	SMPL
PB168135BL - 1	PBW	-586	-0.0629	-	7470A	PPB	23 May 2025 09:29:44	U	SMPL
PB168135BS - 1	LCSW	15368	4.0446	-	7470A	PPB	23 May 2025 09:32:03	U	SMPL
Q2078-03 - 1	WC-A4-04-C	-384	-0.0109	-	7470A	PPB	23 May 2025 09:34:19	U	SMPL
Q2078-07 - 1	WC-A4-05-C	-208	0.0344	-	7470A	PPB	23 May 2025 09:36:38	U	SMPL
Q2078-11 - 1	WC-A1-06A-C	-182	0.0411	-	7470A	PPB	23 May 2025 09:38:54	U	SMPL
Q2078-15 - 1	WC-A1-07A-C	-181	0.0414	-	7470A	PPB	23 May 2025 09:41:11	U	SMPL
Q2078-19 - 1	WC-A4-06-C	-163	0.0460	-	7470A	PPB	23 May 2025 09:43:26	U	SMPL
CCV56 - 1	CCV56	20792	5.4411	-	7470A	PPB	23 May 2025 09:45:41	U	SMPL
CCB56 - 1	CCB56	-478	-0.0351	-	7470A	PPB	23 May 2025 09:47:57	U	SMPL
Q2079-03 - 1	WC-A2-01-C	-266	0.0195	-	7470A	PPB	23 May 2025 09:50:15	U	SMPL
Q2079-07 - 1	WC-A2-02-C	-232	0.0282	-	7470A	PPB	23 May 2025 09:52:30	U	SMPL
Q2079-11 - 1	WC-A2-03-C	-199	0.0367	-	7470A	PPB	23 May 2025 09:54:45	U	SMPL
Q2084-05 - 1	OR-640-COMP-66	-98	0.0627	-	7470A	PPB	23 May 2025 09:57:01	U	SMPL
Q2092-02 - 1	VNJ-202	-97	0.0630	-	7470A	PPB	23 May 2025 09:59:17	U	SMPL
Q2093-02 - 1	SOIL-DRUM	-56	0.0735	-	7470A	PPB	23 May 2025 10:01:33	U	SMPL
Q2095-04 - 1	WCS-TP1	-92	0.0643	-	7470A	PPB	23 May 2025 10:03:50	U	SMPL
Q2095-08 - 1	WCS-TP2	-94	0.0638	-	7470A	PPB	23 May 2025 10:06:06	U	SMPL
Q2095-08DUP - 1	WCS-TP2DUP	-76	0.0684	-	7470A	PPB	23 May 2025 10:08:23	U	SMPL
Q2095-08MS - 1	WCS-TP2MS	16322	4.2902	-	7470A	PPB	23 May 2025 10:10:39	U	SMPL
CCV57 - 1	CCV57	20834	5.4519	-	7470A	PPB	23 May 2025 10:12:56	U	SMPL
CCB57 - 1	CCB57	-499	-0.0405	-	7470A	PPB	23 May 2025 10:15:14	U	SMPL
Q2095-08MSD - 1	WCS-TP2MSD	15075	3.9692	-	7470A	PPB	23 May 2025 10:17:32	U	SMPL
PB168136BL - 1	PBW	-297	0.0115	-	7470A	PPB	23 May 2025 10:19:48	U	SMPL
PB168136BS - 1	LCSW	15547	4.0907	-	7470A	PPB	23 May 2025 10:22:06	U	SMPL
Q2097-02 - 1	ETGI-357	-330	0.0030	-	7470A	PPB	23 May 2025 10:24:22	U	SMPL
Q2097-04 - 1	RBR200044	-125	0.0558	-	7470A	PPB	23 May 2025 10:26:39	U	SMPL
Q2097-06 - 1	RBR251675	-153	0.0486	-	7470A	PPB	23 May 2025 10:28:55	U	SMPL
Q2097-08 - 1	VNJ-244	-166	0.0452	-	7470A	PPB	23 May 2025 10:31:12	U	SMPL
Q2097-10 - 1	ETGI-290	-162	0.0462	-	7470A	PPB	23 May 2025 10:33:29	U	SMPL
Q2097-12 - 1	RT3419	-218	0.0318	-	7470A	PPB	23 May 2025 10:35:46	U	SMPL
Q2097-14 - 1	RBR251372	-191	0.0388	-	7470A	PPB	23 May 2025 10:38:02	U	SMPL
CCV58 - 1	CCV58	20653	5.4053	-	7470A	PPB	23 May 2025 10:40:18	U	SMPL
CCB58 - 1	CCB58	-305	0.0094	-	7470A	PPB	23 May 2025 10:42:34	U	SMPL
Q2097-16 - 1	72-12013	-147	0.0501	-	7470A	PPB	23 May 2025 10:44:51	U	SMPL
Q2097-18 - 1	RT-3888	-115	0.0583	-	7470A	PPB	23 May 2025 10:47:07	U	SMPL
Q2101-04 - 1	TP-1-MHE	-91	0.0645	-	7470A	PPB	23 May 2025 10:49:23	U	SMPL
Q2101-04DUP - 1	TP-1-MHEDUP	-124	0.0560	-	7470A	PPB	23 May 2025 10:51:38	U	SMPL
Q2101-04MS - 1	TP-1-MHEMS	15584	4.1002	-	7470A	PPB	23 May 2025 10:53:54	U	SMPL
Q2101-04MSD - 1	TP-1-MHEMSD	14772	3.8912	-	7470A	PPB	23 May 2025 10:56:10	U	SMPL
CCV59 - 1	CCV59	19955	5.2256	-	7470A	PPB	23 May 2025 10:58:28	U	SMPL
CCB59 - 1	CCB59	-493	-0.0390	-	7470A	PPB	23 May 2025 11:00:47	U	SMPL
PB168092TB - 1	PB168092TB	-184	0.0406	-	7470A	PPB	23 May 2025 11:03:06	U	SMPL
PB168115TB - 1	PB168115TB	-86	0.0658	-	7470A	PPB	23 May 2025 11:05:22	U	SMPL
Q2095-08LX5 - 1		-74	0.0689	-	7470A	PPB	23 May 2025 11:07:37	U	SMPL
Q2095-08A - 1		15423	4.0588	-	7470A	PPB	23 May 2025 11:09:53	U	SMPL
Q2101-04LX5 - 1		-268	0.0190	-	7470A	PPB	23 May 2025 11:12:09	U	SMPL
Q2101-04A - 1		15666	4.1213	-	7470A	PPB	23 May 2025 11:14:26	U	SMPL
CCV60 - 1	CCV60	18813	4.9316	-	7470A	PPB	23 May 2025 11:23:20	U	SMPL
CCB60 - 1	CCB60	-257	0.0218	-	7470A	PPB	23 May 2025 11:25:35	U	SMPL