

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2078
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2078 **SAS No.:** Q2078
Matrix: Water **Sample ID:** Q2095-08 **Client ID:** WCS-TP2DUP
Percent Solids for Sample: NA **Duplicate ID** Q2095-08DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1340		1410		5		P
Cadmium	ug/L	20	20.7	J	30.0	U	200.0		P
Chromium	ug/L	20	16.5	J	50.0	U	200.0		P
Copper	ug/L	20	57.2	J	39.7	J	36		P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	51.4	J	27.0	J	62		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	298		226		27		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2078
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2078 **SAS No.:** Q2078
Matrix: Water **Sample ID:** Q2095-08MS **Client ID:** WCS-TP2MSD
Percent Solids for Sample: NA **Duplicate ID** Q2095-08MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3970		3960		0		P
Barium	ug/L	20	2160		2590		18		P
Cadmium	ug/L	20	975		972		0		P
Chromium	ug/L	20	2030		2070		2		P
Copper	ug/L	20	1510		1500		1		P
Lead	ug/L	20	4650		4620		1		P
Mercury	ug/L	20	42.9		39.7		8		CV
Nickel	ug/L	20	2480		2500		1		P
Selenium	ug/L	20	9810		9810		0		P
Silver	ug/L	20	360		356		1		P
Zinc	ug/L	20	1270		1240		2		P