

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

Continuing Calib Date: 05/21/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:04 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.66	5.66	5.56	5.76	0.00
Aroclor-1248-2 (2)	5.93	5.93	5.83	6.03	0.00
Aroclor-1248-3 (3)	6.13	6.13	6.03	6.23	0.00
Aroclor-1248-4 (4)	6.53	6.53	6.43	6.63	0.00
Aroclor-1248-5 (5)	6.57	6.57	6.47	6.67	0.00
Tetrachloro-m-xylene	4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl	10.21	10.22	10.12	10.32	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

Continuing Calib Date: 05/21/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:04 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1248-2	(2)	5.12	5.12	5.02	5.22	0.00
Aroclor-1248-3	(3)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-4	(4)	5.33	5.33	5.23	5.43	0.00
Aroclor-1248-5	(5)	5.72	5.73	5.63	5.83	0.01
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.82	8.83	8.73	8.93	0.01