

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072239.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.655	5.552	5.752	474.270	500.000	-5.1
Aroclor-1016-2	5.677	5.574	5.774	516.170	500.000	3.2
Aroclor-1016-3	5.739	5.636	5.836	505.350	500.000	1.1
Aroclor-1016-4	5.836	5.734	5.934	523.580	500.000	4.7
Aroclor-1016-5	6.129	6.027	6.227	530.670	500.000	6.1
Aroclor-1260-1	7.247	7.147	7.347	526.120	500.000	5.2
Aroclor-1260-2	7.501	7.401	7.601	514.240	500.000	2.8
Aroclor-1260-3	7.860	7.760	7.960	530.610	500.000	6.1
Aroclor-1260-4	8.084	7.985	8.185	534.440	500.000	6.9
Aroclor-1260-5	8.403	8.303	8.503	516.550	500.000	3.3
Decachlorobiphenyl	10.209	10.113	10.313	51.950	50.000	3.9
Tetrachloro-m-xylene	4.502	4.396	4.596	52.210	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072239.D Time Analyzed: 09:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	516.180	500.000	3.2
Aroclor-1016-2	4.896	4.795	4.995	525.670	500.000	5.1
Aroclor-1016-3	5.072	4.973	5.173	536.380	500.000	7.3
Aroclor-1016-4	5.114	5.015	5.215	538.420	500.000	7.7
Aroclor-1016-5	5.328	5.229	5.429	522.650	500.000	4.5
Aroclor-1260-1	6.361	6.264	6.464	551.640	500.000	10.3
Aroclor-1260-2	6.550	6.453	6.653	522.460	500.000	4.5
Aroclor-1260-3	6.702	6.606	6.806	548.610	500.000	9.7
Aroclor-1260-4	7.173	7.077	7.277	544.680	500.000	8.9
Aroclor-1260-5	7.415	7.319	7.519	535.160	500.000	7.0
Decachlorobiphenyl	8.817	8.723	8.923	54.600	50.000	9.2
Tetrachloro-m-xylene	3.794	3.691	3.891	51.300	50.000	2.6