

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072254.D Time Analyzed: 16:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.657	5.552	5.752	480.150	500.000	-4.0
Aroclor-1016-2	5.679	5.574	5.774	511.690	500.000	2.3
Aroclor-1016-3	5.741	5.636	5.836	504.550	500.000	0.9
Aroclor-1016-4	5.839	5.734	5.934	519.060	500.000	3.8
Aroclor-1016-5	6.131	6.027	6.227	518.870	500.000	3.8
Aroclor-1260-1	7.250	7.147	7.347	524.010	500.000	4.8
Aroclor-1260-2	7.503	7.401	7.601	506.070	500.000	1.2
Aroclor-1260-3	7.862	7.760	7.960	521.850	500.000	4.4
Aroclor-1260-4	8.087	7.985	8.185	522.920	500.000	4.6
Aroclor-1260-5	8.406	8.303	8.503	516.620	500.000	3.3
Decachlorobiphenyl	10.212	10.113	10.313	51.990	50.000	4.0
Tetrachloro-m-xylene	4.504	4.396	4.596	51.680	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072254.D Time Analyzed: 16:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.880	4.777	4.977	513.030	500.000	2.6
Aroclor-1016-2	4.898	4.795	4.995	516.100	500.000	3.2
Aroclor-1016-3	5.074	4.973	5.173	529.340	500.000	5.9
Aroclor-1016-4	5.116	5.015	5.215	529.500	500.000	5.9
Aroclor-1016-5	5.330	5.229	5.429	553.920	500.000	10.8
Aroclor-1260-1	6.363	6.264	6.464	547.180	500.000	9.4
Aroclor-1260-2	6.552	6.453	6.653	517.120	500.000	3.4
Aroclor-1260-3	6.705	6.606	6.806	548.310	500.000	9.7
Aroclor-1260-4	7.176	7.077	7.277	548.580	500.000	9.7
Aroclor-1260-5	7.417	7.319	7.519	544.930	500.000	9.0
Decachlorobiphenyl	8.820	8.723	8.923	56.140	50.000	12.3
Tetrachloro-m-xylene	3.797	3.691	3.891	53.300	50.000	6.6