

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072269.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.653	5.552	5.752	474.380	500.000	-5.1
Aroclor-1016-2	5.674	5.574	5.774	504.990	500.000	1.0
Aroclor-1016-3	5.737	5.636	5.836	490.580	500.000	-1.9
Aroclor-1016-4	5.834	5.734	5.934	501.930	500.000	0.4
Aroclor-1016-5	6.127	6.027	6.227	506.390	500.000	1.3
Aroclor-1260-1	7.245	7.147	7.347	498.780	500.000	-0.2
Aroclor-1260-2	7.499	7.401	7.601	477.400	500.000	-4.5
Aroclor-1260-3	7.857	7.760	7.960	486.380	500.000	-2.7
Aroclor-1260-4	8.082	7.985	8.185	483.510	500.000	-3.3
Aroclor-1260-5	8.401	8.303	8.503	498.490	500.000	-0.3
Decachlorobiphenyl	10.207	10.113	10.313	50.420	50.000	0.8
Tetrachloro-m-xylene	4.500	4.396	4.596	52.020	50.000	4.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072269.D Time Analyzed: 22:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	497.510	500.000	-0.5
Aroclor-1016-2	4.896	4.795	4.995	506.510	500.000	1.3
Aroclor-1016-3	5.073	4.973	5.173	523.110	500.000	4.6
Aroclor-1016-4	5.115	5.015	5.215	521.750	500.000	4.4
Aroclor-1016-5	5.329	5.229	5.429	503.860	500.000	0.8
Aroclor-1260-1	6.362	6.264	6.464	507.640	500.000	1.5
Aroclor-1260-2	6.551	6.453	6.653	486.840	500.000	-2.6
Aroclor-1260-3	6.703	6.606	6.806	535.950	500.000	7.2
Aroclor-1260-4	7.174	7.077	7.277	525.860	500.000	5.2
Aroclor-1260-5	7.416	7.319	7.519	538.710	500.000	7.7
Decachlorobiphenyl	8.819	8.723	8.923	53.420	50.000	6.8
Tetrachloro-m-xylene	3.795	3.691	3.891	52.290	50.000	4.6