

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344, PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359, PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072162.D	19 May 2025 09:09	YP\AJ	Ok
2	I.BLK	PP072163.D	19 May 2025 09:26	YP\AJ	Ok
3	AR1660ICC1000	PP072164.D	19 May 2025 09:42	YP\AJ	Ok
4	AR1660ICC750	PP072165.D	19 May 2025 09:59	YP\AJ	Ok
5	AR1660ICC500	PP072166.D	19 May 2025 10:15	YP\AJ	Ok
6	AR1660ICC250	PP072167.D	19 May 2025 10:32	YP\AJ	Ok
7	AR1660ICC050	PP072168.D	19 May 2025 10:48	YP\AJ	Ok,M
8	AR1221ICC500	PP072169.D	19 May 2025 11:05	YP\AJ	Ok
9	AR1232ICC1000	PP072170.D	19 May 2025 11:21	YP\AJ	Ok
10	AR1232ICC750	PP072171.D	19 May 2025 11:38	YP\AJ	Ok
11	AR1232ICC500	PP072172.D	19 May 2025 11:54	YP\AJ	Ok
12	AR1232ICC250	PP072173.D	19 May 2025 12:10	YP\AJ	Ok
13	AR1232ICC050	PP072174.D	19 May 2025 12:27	YP\AJ	Ok,M
14	AR1242ICC1000	PP072175.D	19 May 2025 12:43	YP\AJ	Ok
15	AR1242ICC750	PP072176.D	19 May 2025 12:59	YP\AJ	Ok
16	AR1242ICC500	PP072177.D	19 May 2025 13:15	YP\AJ	Ok
17	AR1242ICC250	PP072178.D	19 May 2025 13:32	YP\AJ	Ok
18	AR1242ICC050	PP072179.D	19 May 2025 13:48	YP\AJ	Ok,M
19	AR1248ICC1000	PP072180.D	19 May 2025 14:04	YP\AJ	Ok
20	AR1248ICC750	PP072181.D	19 May 2025 14:20	YP\AJ	Ok
21	AR1248ICC500	PP072182.D	19 May 2025 14:37	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1248ICC250	PP072183.D	19 May 2025 14:53	YP\AJ	Ok
23	AR1248ICC050	PP072184.D	19 May 2025 15:25	YP\AJ	Ok,M
24	AR1254ICC1000	PP072185.D	19 May 2025 15:42	YP\AJ	Ok
25	AR1254ICC750	PP072186.D	19 May 2025 15:58	YP\AJ	Ok
26	AR1254ICC500	PP072187.D	19 May 2025 16:14	YP\AJ	Ok
27	AR1254ICC250	PP072188.D	19 May 2025 16:31	YP\AJ	Ok
28	AR1254ICC050	PP072189.D	19 May 2025 16:47	YP\AJ	Ok,M
29	AR1262ICC500	PP072190.D	19 May 2025 17:03	YP\AJ	Ok
30	AR1268ICC1000	PP072191.D	19 May 2025 17:20	YP\AJ	Ok
31	AR1268ICC750	PP072192.D	19 May 2025 17:36	YP\AJ	Ok
32	AR1268ICC500	PP072193.D	19 May 2025 17:53	YP\AJ	Ok
33	AR1268ICC250	PP072194.D	19 May 2025 18:09	YP\AJ	Ok
34	AR1268ICC050	PP072195.D	19 May 2025 18:25	YP\AJ	Ok,M
35	PP051925ICV500	PP072196.D	19 May 2025 18:42	YP\AJ	Ok
36	AR1232ICV500	PP072197.D	19 May 2025 19:14	YP\AJ	Ok
37	AR1242ICV500	PP072198.D	19 May 2025 19:47	YP\AJ	Ok
38	AR1248ICV500	PP072199.D	19 May 2025 20:19	YP\AJ	Ok
39	AR1254ICV500	PP072200.D	19 May 2025 20:52	YP\AJ	Ok
40	AR1268ICV500	PP072201.D	19 May 2025 21:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP052125

Review By	yogesh	Review On	5/21/2025 10:53:17 AM
Supervise By		Supervise On	
SubDirectory	PP052125	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072238.D	21 May 2025 09:00	YP\AJ	Ok
2	AR1660CCC500	PP072239.D	21 May 2025 09:16	YP\AJ	Ok
3	AR1242CCC500	PP072240.D	21 May 2025 09:47	YP\AJ	Ok
4	AR1248CCC500	PP072241.D	21 May 2025 10:04	YP\AJ	Ok
5	AR1254CCC500	PP072242.D	21 May 2025 10:20	YP\AJ	Ok,NS
6	I.BLK	PP072243.D	21 May 2025 10:36	YP\AJ	Ok
7	PB168081BL	PP072244.D	21 May 2025 10:53	YP\AJ	Ok
8	Q2067-01	PP072245.D	21 May 2025 11:09	YP\AJ	Ok,NS
9	PB168099BL	PP072246.D	21 May 2025 13:33	YP\AJ	Ok
10	PB168099BS	PP072247.D	21 May 2025 13:49	YP\AJ	Ok
11	Q2078-02	PP072248.D	21 May 2025 14:06	YP\AJ	Ok,NS
12	Q2078-06	PP072249.D	21 May 2025 14:22	YP\AJ	Ok,NS
13	Q2078-10	PP072250.D	21 May 2025 14:39	YP\AJ	Ok,NS
14	Q2078-14	PP072251.D	21 May 2025 14:55	YP\AJ	Ok,NS
15	Q2078-18	PP072252.D	21 May 2025 15:12	YP\AJ	Ok,NS
16	Q2079-02	PP072253.D	21 May 2025 15:28	YP\AJ	Ok,NS
17	AR1660CCC500	PP072254.D	21 May 2025 16:51	YP\AJ	Ok
18	AR1242CCC500	PP072255.D	21 May 2025 17:40	YP\AJ	Ok
19	AR1248CCC500	PP072256.D	21 May 2025 17:56	YP\AJ	Ok
20	AR1254CCC500	PP072257.D	21 May 2025 18:12	YP\AJ	Ok,NR
21	I.BLK	PP072258.D	21 May 2025 18:29	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP052125

Review By	yogesh	Review On	5/21/2025 10:53:17 AM
Supervise By		Supervise On	
SubDirectory	PP052125	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2079-06	PP072259.D	21 May 2025 18:45	YPIAJ	Ok,NS
23	Q2079-10	PP072260.D	21 May 2025 19:01	YPIAJ	Ok,NS
24	Q2084-01	PP072261.D	21 May 2025 19:17	YPIAJ	Ok,NS
25	Q2084-01MS	PP072262.D	21 May 2025 19:34	YPIAJ	Ok,NS
26	Q2084-01MSD	PP072263.D	21 May 2025 19:50	YPIAJ	Ok,NS
27	Q2087-01	PP072264.D	21 May 2025 20:06	YPIAJ	Ok,NS
28	Q2087-03	PP072265.D	21 May 2025 20:22	YPIAJ	Ok,NS
29	Q2097-01	PP072266.D	21 May 2025 20:39	YPIAJ	Ok,NS
30	Q2097-03	PP072267.D	21 May 2025 20:55	YPIAJ	Ok
31	Q2097-05	PP072268.D	21 May 2025 21:11	YPIAJ	Ok
32	AR1660CCC500	PP072269.D	21 May 2025 22:17	YPIAJ	Ok
33	AR1242CCC500	PP072270.D	21 May 2025 23:05	YPIAJ	Ok
34	AR1248CCC500	PP072271.D	21 May 2025 23:22	YPIAJ	Ok
35	AR1254CCC500	PP072272.D	21 May 2025 23:38	YPIAJ	Ok,NR
36	I.BLK	PP072273.D	21 May 2025 23:54	YPIAJ	Ok
37	PB168118BL	PP072274.D	22 May 2025 00:11	YPIAJ	Ok
38	PB168118BS	PP072275.D	22 May 2025 00:27	YPIAJ	Ok
39	Q2103-01	PP072276.D	22 May 2025 00:43	YPIAJ	Ok
40	Q2103-02	PP072277.D	22 May 2025 01:00	YPIAJ	Ok
41	Q2103-03	PP072278.D	22 May 2025 01:16	YPIAJ	Ok,NS
42	Q2103-04	PP072279.D	22 May 2025 01:32	YPIAJ	Ok
43	AR1660CCC500	PP072280.D	22 May 2025 02:37	YPIAJ	Ok
44	AR1242CCC500	PP072281.D	22 May 2025 03:27	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP052125

Review By	yogesh	Review On	5/21/2025 10:53:17 AM
Supervise By		Supervise On	
SubDirectory	PP052125	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	AR1248CCC500	PP072282.D	22 May 2025 03:43	YP\AJ	Ok
46	AR1254CCC500	PP072283.D	22 May 2025 03:59	YP\AJ	Ok
47	I.BLK	PP072284.D	22 May 2025 04:16	YP\AJ	Ok

M : Manual Integration