

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2078 **SAS No.:** Q2078 **SDG NO.:** Q2078
Instrument ID: ECD_S **Calibration Date(s):** 05/12/2025 05/12/2025
Calibration Times: 12:30 14:06

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS030125.D</u>		CF 500 = <u>PS030126.D</u>			
CF 750 = <u>PS030127.D</u>		CF 1000 = <u>PS030128.D</u>		CF 1500 = <u>PS030129.D</u>			
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	19321200000	16303400000	15619800000	15128700000	14606800000	16196000000	11
2,4-D	4048870000	3291910000	3113030000	3014870000	2926010000	3278940000	14
2,4-DCAA	3540840000	2868830000	2702520000	2609290000	2517560000	2847810000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2078 **SAS No.:** Q2078 **SDG NO.:** Q2078
Instrument ID: ECD_S **Calibration Date(s):** 05/12/2025 05/12/2025
Calibration Times: 12:30 14:06

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS030125.D</u>		CF 500 = <u>PS030126.D</u>			
CF 750 = <u>PS030127.D</u>		CF 1000 = <u>PS030128.D</u>		CF 1500 = <u>PS030129.D</u>			
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	10913400000	9868000000	9697110000	9517670000	9255920000	9850430000	6
2,4-D	1480680000	1290430000	1248060000	1225910000	1207420000	1290500000	9
2,4-DCAA	922885000	799420000	773490000	758361000	746863000	800204000	9