

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ENTA05

**Lab Code:** CHEM      **Case No.:** Q2078      **SAS No.:** Q2078      **SDG NO.:** Q2078

**Continuing Calib Date:** 05/23/2025      **Initial Calibration Date(s):** 05/12/2025      05/12/2025

**Continuing Calib Time:** 00:54      **Initial Calibration Time(s):** 12:30      14:06

**GC Column:** RTX-CLP      **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 6.93       | 6.93      | 6.83      | 7.03 | 0.00       |
| 2,4-D            | 8.00       | 8.01      | 7.91      | 8.11 | 0.01       |
| 2,4,5-TP(Silvex) | 8.84       | 8.85      | 8.75      | 8.95 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ENTA05

**Lab Code:** CHEM      **Case No.:** Q2078      **SAS No.:** Q2078      **SDG NO.:** Q2078

**Continuing Calib Date:** 05/23/2025      **Initial Calibration Date(s):** 05/12/2025      05/12/2025

**Continuing Calib Time:** 00:54      **Initial Calibration Time(s):** 12:30      14:06

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.45       | 7.45      | 7.35      | 7.55 | 0.00       |
| 2,4-D            | 8.63       | 8.64      | 8.54      | 8.74 | 0.01       |
| 2,4,5-TP(Silvex) | 9.51       | 9.52      | 9.42      | 9.62 | 0.01       |