

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030348.D Time Analyzed: 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.845	8.750	8.950	721.210	712.500	1.2
2,4-D	8.005	7.910	8.110	708.190	705.000	0.5
2,4-DCAA	6.929	6.834	7.034	748.510	750.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030348.D Time Analyzed: 16:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.513	9.423	9.623	719.420	712.500	1.0
2,4-D	8.635	8.544	8.744	704.970	705.000	0.0
2,4-DCAA	7.447	7.354	7.554	731.090	750.000	-2.5