

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2078

Analytical Method: 8151A

Client: ENTACT

Datafile : PS030344.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168120BS (Column 1)	2,4-D	5	5.20	ug/L	104				70 (83)	130 (130)	
	2,4,5-TP(Silvex)	5	5.20	ug/L	104				70 (78)	130 (127)	
PB168120BS (Column 2)	2,4-D	5	5.10	ug/L	102				70 (83)	130 (130)	
	2,4,5-TP(Silvex)	5	5.10	ug/L	102				70 (78)	130 (127)	