

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168121BS

Contract: ENTA05

Lab Code: CHEM

Case No.: Q2078

SAS No.: Q2078

SDG NO.: Q2078

Lab Sample ID: PB168121BS

Date(s) Analyzed: 05/22/2025

05/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.53	7.48	7.58	0.44	8.1
	2	6.75	6.70	6.80	0.40	
gamma-BHC (Lindane)	1	4.36	4.31	4.41	0.57	3.4
	2	3.73	3.68	3.78	0.55	
Heptachlor	1	4.96	4.91	5.01	0.54	4.8
	2	4.08	4.03	4.13	0.51	
Heptachlor epoxide	1	5.72	5.67	5.77	0.58	4.8
	2	4.87	4.82	4.92	0.56	
Endrin	1	6.60	6.55	6.65	0.47	3.6
	2	5.78	5.73	5.83	0.45	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PL-HRH-COMP-01MS

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078

SAS No.: Q2078 SDG NO.: Q2078

Lab Sample ID: Q2080-06MS

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.52	7.47	7.57	4.20	2.4
	2	6.75	6.70	6.80	4.10	
gamma-BHC (Lindane)	1	4.36	4.31	4.41	5.10	0
	2	3.73	3.68	3.78	5.10	
Heptachlor	1	4.95	4.90	5.00	4.80	2.1
	2	4.08	4.03	4.13	4.70	
Heptachlor epoxide	1	5.72	5.67	5.77	5.00	2
	2	4.87	4.82	4.92	5.10	
Endrin	1	6.60	6.55	6.65	4.50	0
	2	5.78	5.73	5.83	4.50	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PL-HRH-COMP-01MSD

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078

SAS No.: Q2078 SDG NO.: Q2078

Lab Sample ID: Q2080-06MSD

Date(s) Analyzed: 05/22/2025 05/22/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.60	6.55	6.65	4.60	2.2
	2	5.78	5.73	5.83	4.50	
Methoxychlor	1	7.52	7.47	7.57	4.30	7.2
	2	6.75	6.70	6.80	4.00	
gamma-BHC (Lindane)	1	4.36	4.31	4.41	4.90	2
	2	3.73	3.68	3.78	5.00	
Heptachlor	1	4.95	4.90	5.00	4.80	2.1
	2	4.08	4.03	4.13	4.70	
Heptachlor epoxide	1	5.72	5.67	5.77	5.10	2
	2	4.87	4.82	4.92	5.00	