

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095756.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.103	9.003	9.203	50.300	50.000	0.6
Endrin	6.603	6.502	6.702	45.920	50.000	-8.2
gamma-BHC (Lindane)	4.356	4.257	4.457	53.810	50.000	7.6
Heptachlor	4.954	4.855	5.055	51.070	50.000	2.1
Heptachlor epoxide	5.718	5.618	5.818	54.020	50.000	8.0
Methoxychlor	7.524	7.423	7.623	44.350	50.000	-11.3
Tetrachloro-m-xylene	3.573	3.473	3.673	53.640	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095756.D Time Analyzed: 09:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.062	7.964	8.164	53.040	50.000	6.1
Endrin	5.781	5.683	5.883	44.550	50.000	-10.9
gamma-BHC (Lindane)	3.729	3.630	3.830	52.870	50.000	5.7
Heptachlor	4.082	3.983	4.183	49.500	50.000	-1.0
Heptachlor epoxide	4.869	4.771	4.971	52.510	50.000	5.0
Methoxychlor	6.748	6.651	6.851	42.280	50.000	-15.4
Tetrachloro-m-xylene	2.885	2.786	2.986	51.930	50.000	3.9