

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095777.D Time Analyzed: 17:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.104	9.003	9.203	52.180	50.000	4.4
Endrin	6.602	6.502	6.702	45.900	50.000	-8.2
gamma-BHC (Lindane)	4.356	4.257	4.457	55.570	50.000	11.1
Heptachlor	4.954	4.855	5.055	52.110	50.000	4.2
Heptachlor epoxide	5.718	5.618	5.818	56.230	50.000	12.5
Methoxychlor	7.523	7.423	7.623	43.460	50.000	-13.1
Tetrachloro-m-xylene	3.573	3.473	3.673	55.450	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095777.D Time Analyzed: 17:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.063	7.964	8.164	55.650	50.000	11.3
Endrin	5.782	5.683	5.883	44.970	50.000	-10.1
gamma-BHC (Lindane)	3.730	3.630	3.830	54.810	50.000	9.6
Heptachlor	4.083	3.983	4.183	50.640	50.000	1.3
Heptachlor epoxide	4.870	4.771	4.971	54.700	50.000	9.4
Methoxychlor	6.750	6.651	6.851	41.650	50.000	-16.7
Tetrachloro-m-xylene	2.886	2.786	2.986	53.640	50.000	7.3