

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG No.: Q2078

Instrument ID: MSVOA_N Calibration Date/Time: 05/21/2025 11:10

Lab File ID: VN086722.D Init. Calib. Date(s): 05/16/2025 05/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:31 18:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.733	0.782		6.68	20
1,1-Dichloroethene	0.558	0.579		3.76	20
2-Butanone	0.399	0.400		0.25	20
Carbon Tetrachloride	0.433	0.457		5.54	20
Chloroform	1.119	1.149		2.68	20
Benzene	1.433	1.494		4.26	20
1,2-Dichloroethane	0.446	0.453		1.57	20
Trichloroethene	0.346	0.363		4.91	20
Tetrachloroethene	0.383	0.414		8.09	20
Chlorobenzene	1.096	1.139	0.3	3.92	20
1,2-Dichloroethane-d4	0.691	0.701		1.45	20
Dibromofluoromethane	0.298	0.307		3.02	20
Toluene-d8	1.234	1.283		3.97	20
4-Bromofluorobenzene	0.436	0.470		7.57	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.