

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN051625

Review By	John Carlone	Review On	5/19/2025 9:34:08 AM
Supervise By	Maresh Dadoda	Supervise On	5/19/2025 2:40:58 PM
SubDirectory	VN051625	HP Acquire Method	HP Processing Method 624N042325W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133946,VP133952
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133947,VP133948 VP133981

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086660.D	16 May 2025 09:17		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN086661.D	16 May 2025 11:37	pH#Lot#V12668	JC\MD	Ok,M
3	VN0516WBS01	VN0516WBS01	VN086662.D	16 May 2025 12:13		JC\MD	Ok,M
4	VN0516WBSD01	VN0516WBSD01	VN086663.D	16 May 2025 12:46		JC\MD	Ok,M
5	VN0516WBL01	VN0516WBL01	VN086664.D	16 May 2025 13:20		JC\MD	Ok
6	Q2006-01	EFF-WW	VN086665.D	16 May 2025 13:54	vial A pH<2	JC\MD	Ok
7	IBLK	IBLK	VN086666.D	16 May 2025 14:46		JC\MD	Ok
8	VSTDCCC020	VSTDCCC020EC	VN086667.D	16 May 2025 15:09		JC\MD	Ok,M
9	BFB	BFB	VN086668.D	16 May 2025 15:33		JC\MD	Ok
10	VSTDICC001	VSTDICC001	VN086669.D	16 May 2025 16:31		JC\MD	Ok,M
11	VSTDICC005	VSTDICC005	VN086670.D	16 May 2025 16:55		JC\MD	Ok,M
12	VSTDICC010	VSTDICC010	VN086671.D	16 May 2025 17:19		JC\MD	Ok,M
13	VSTDICC020	VSTDICC020	VN086672.D	16 May 2025 17:43		JC\MD	Ok,M
14	VSTDICCC050	VSTDICCC050	VN086673.D	16 May 2025 18:08		JC\MD	Ok,M
15	VSTDICC100	VSTDICC100	VN086674.D	16 May 2025 18:32		JC\MD	Ok,M
16	IBLK	IBLK	VN086675.D	16 May 2025 18:56		JC\MD	Ok
17	VSTDICV050	ICVVN051625	VN086676.D	16 May 2025 19:20		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052125

Review By	John Carlone	Review On	5/22/2025 8:01:52 AM
Supervise By	Maresh Dadoda	Supervise On	5/22/2025 8:24:13 AM
SubDirectory	VN052125	HP Acquire Method	HP Processing Method 82N051625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133982
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133983,VP133984

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086721.D	21 May 2025 10:37		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086722.D	21 May 2025 11:10	pH#Lot#V12668	JC\MD	Ok,M
3	VN0521MBL01	VN0521MBL01	VN086723.D	21 May 2025 11:44		JC\MD	Ok
4	VN0521WBL01	VN0521WBL01	VN086724.D	21 May 2025 12:08		JC\MD	Ok
5	VN0521MBS01	VN0521MBS01	VN086725.D	21 May 2025 12:32		JC\MD	Ok,M
6	VN0521WBS01	VN0521WBS01	VN086726.D	21 May 2025 13:05		JC\MD	Ok,M
7	Q2068-01	LAW-25-0075	VN086727.D	21 May 2025 13:29	run straight MEOH	JC\MD	Not Ok
8	VN0521WBSD01	VN0521WBSD01	VN086728.D	21 May 2025 13:53		JC\MD	Ok,M
9	Q2068-01	LAW-25-0075	VN086729.D	21 May 2025 14:17		JC\MD	Ok
10	PB168093TB	PB168093TB	VN086730.D	21 May 2025 14:41		JC\MD	Ok
11	Q2078-05	WC-A4-05-G	VN086731.D	21 May 2025 15:05	vial A pH#5.0 Surrogate fail;Need 20X	JC\MD	Dilution
12	Q2078-09	WC-A1-06A-G	VN086732.D	21 May 2025 15:29	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
13	Q2078-13	WC-A1-07A-G	VN086733.D	21 May 2025 15:53	vial A pH#5.0	JC\MD	Ok
14	Q2078-17	WC-A4-06-G	VN086734.D	21 May 2025 16:18	vial A pH#5.0	JC\MD	Ok
15	Q2079-01	WC-A2-01-G	VN086735.D	21 May 2025 16:42	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
16	Q2079-05	WC-A2-02-G	VN086736.D	21 May 2025 17:06	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
17	Q2079-09	WC-A2-03-G	VN086737.D	21 May 2025 17:30	vial A pH#5.0	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052125

Review By	John Carlone	Review On	5/22/2025 8:01:52 AM
Supervise By	Maresh Dadoda	Supervise On	5/22/2025 8:24:13 AM
SubDirectory	VN052125	HP Acquire Method	HP Processing Method 82N051625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133982		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133983,VP133984		

18	Q2084-05	OR-640-COMP-66	VN086738.D	21 May 2025 17:54	vial A pH#5.0	JC\MD	Ok
19	Q2092-02	VNJ-202	VN086739.D	21 May 2025 18:18	vial A pH#5.0	JC\MD	Ok
20	Q2093-02	SOIL-DRUM	VN086740.D	21 May 2025 18:42	vial A pH#5.0	JC\MD	Ok
21	Q2095-04	WCS-TP1	VN086741.D	21 May 2025 19:06	vial A pH#5.0	JC\MD	Ok
22	Q2095-08	WCS-TP2	VN086742.D	21 May 2025 19:31	vial A pH#5.0 RT Shifted	JC\MD	Ok,M
23	Q2078-01	WC-A4-04-G	VN086743.D	21 May 2025 19:54	vial A pH#5.0 Surrogate Fail	JC\MD	ReRun
24	Q2032-01ME	TP-11ME	VN086744.D	21 May 2025 20:18		JC\MD	Ok
25	Q2075-06	SS-MW1-11.5	VN086745.D	21 May 2025 20:42	Need Soil Run	JC\MD	Not Ok
26	Q2071-19DL	L2-WC-6-VOC DL	VN086746.D	21 May 2025 21:06		JC\MD	Ok
27	VN0521MBS01	VN0521MBS01	VN086747.D	21 May 2025 21:30	Recovery failed	JC\MD	Not Ok
28	VSTDCCC050	VSTDCCC050EC	VN086748.D	21 May 2025 21:54		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN052225

Review By	Review On
Supervise By	Supervise On
SubDirectory	VN052225
HP Acquire Method	HP Processing Method
	82N051625W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133988
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133989,VP133990

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086749.D	22 May 2025 07:48		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086750.D	22 May 2025 08:21		JC\MD	Ok,NR
3	VN0522MBL01	VN0522MBL01	VN086751.D	22 May 2025 08:55		JC\MD	Ok
4	VN0522WBL01	VN0522WBL01	VN086752.D	22 May 2025 09:19		JC\MD	Ok
5	VN0522WBS01	VN0522WBS01	VN086753.D	22 May 2025 09:43		JC\MD	Ok,NR
6	VN0522MBS01	VN0522MBS01	VN086754.D	22 May 2025 10:18		JC\MD	Ok,NR
7	Q2078-05DL	WC-A4-05-GDL	VN086755.D	22 May 2025 10:42		JC\MD	Ok
8	Q2078-01RE	WC-A4-04-GRE	VN086756.D	22 May 2025 11:06	Surrogate Fail	JC\MD	Confirms
9	Q2078-09RE	WC-A1-06A-GRE	VN086757.D	22 May 2025 11:31	Surrogate Fail	JC\MD	Confirms
10	Q2079-01RE	WC-A2-01-GRE	VN086758.D	22 May 2025 11:55	Surrogate Fail	JC\MD	Confirms
11	Q2079-05RE	WC-A2-02-GRE	VN086759.D	22 May 2025 12:19	Surrogate Fail	JC\MD	Confirms
12	VN0522WBSD01	VN0522WBSD01	VN086760.D	22 May 2025 12:43		JC\MD	Ok,NR
13	VN0522MBSD01	VN0522MBSD01	VN086761.D	22 May 2025 13:07		JC\MD	Ok,NR
14	Q2071-19	L2-WC-6-VOC	VN086762.D	22 May 2025 13:31	Need 20X	JC\MD	Dilution
15	IBLK	IBLK	VN086763.D	22 May 2025 13:55		JC\MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VN086764.D	22 May 2025 14:43		JC\MD	Ok,NR

M : Manual Integration