

Surrogate Summary

SDG No.: **Q2078**

Client: **ENTACT**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2078-01	WC-A4-04-G	1,2-Dichloroethane-d4	50	44.1	88	70 (74)	130 (125)
		Dibromofluoromethane	50	35.5	71	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.7	109	70 (77)	130 (121)
Q2078-01RE	WC-A4-04-GRE	1,2-Dichloroethane-d4	50	44.7	89	70 (74)	130 (125)
		Dibromofluoromethane	50	35.0	70	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108	70 (77)	130 (121)
Q2078-05	WC-A4-05-G	1,2-Dichloroethane-d4	50	47.1	94	70 (74)	130 (125)
		Dibromofluoromethane	50	33.2	66 *	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.3	105	70 (77)	130 (121)
Q2078-05DL	WC-A4-05-GDL	1,2-Dichloroethane-d4	50	45.4	91	70 (74)	130 (125)
		Dibromofluoromethane	50	50.8	102	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108	70 (77)	130 (121)
Q2078-09	WC-A1-06A-G	1,2-Dichloroethane-d4	50	44.9	90	70 (74)	130 (125)
		Dibromofluoromethane	50	34.6	69 *	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.1	112	70 (77)	130 (121)
Q2078-09RE	WC-A1-06A-GRE	1,2-Dichloroethane-d4	50	43.7	87	70 (74)	130 (125)
		Dibromofluoromethane	50	35.2	70	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.3	107	70 (77)	130 (121)
Q2078-13	WC-A1-07A-G	1,2-Dichloroethane-d4	50	45.0	90	70 (74)	130 (125)
		Dibromofluoromethane	50	40.6	81	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.9	108	70 (77)	130 (121)
Q2078-17	WC-A4-06-G	1,2-Dichloroethane-d4	50	43.8	88	70 (74)	130 (125)
		Dibromofluoromethane	50	39.5	79	70 (75)	130 (124)
		Toluene-d8	50	52.2	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.8	114	70 (77)	130 (121)
VN0521WBL01	VN0521WBL01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	53.5	107	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108	70 (77)	130 (121)
VN0521WBS01	VN0521WBS01	1,2-Dichloroethane-d4	50	53.9	108	70 (74)	130 (125)
		Dibromofluoromethane	50	53.2	106	70 (75)	130 (124)
		Toluene-d8	50	53.1	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.8	108	70 (77)	130 (121)
VN0521WBSD0	VN0521WBSD01	1,2-Dichloroethane-d4	50	50.0	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)
VN0522WBL01	VN0522WBL01	1,2-Dichloroethane-d4	50	44.2	88	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VN0522WBS01	VN0522WBS01	1,2-Dichloroethane-d4	50	42.1	84	70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100	70 (75)	130 (124)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2078

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0522WBS01	VN0522WBS01	Toluene-d8	50	49.3	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97	70 (77)	130 (121)

