

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078
 Lab File ID: VN086722.D Date Analyzed: 05/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	144979	13.788				
UPPER LIMIT	289958	14.288				
LOWER LIMIT	72489.5	13.288				
EPA SAMPLE NO.						
WC-A4-04-G	207602	13.79				
WC-A4-05-G	186063	13.79				
WC-A1-06A-G	205371	13.79				
WC-A1-07A-G	203184	13.79				
WC-A4-06-G	206737	13.79				
VN0521WBL01	149145	13.79				
VN0521WBS01	127830	13.79				
VN0521WBSD01	140932	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.