

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2078 SAS No.: Q2078 SDG NO.: Q2078
 Lab File ID: VN086750.D Date Analyzed: 05/22/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	188336	13.788				
UPPER LIMIT	376672	14.288				
LOWER LIMIT	94168	13.288				
EPA SAMPLE NO.						
WC-A4-04-GRE	202121	13.79				
WC-A4-05-GDL	193941	13.79				
WC-A1-06A-GRE	195863	13.79				
VN0522WBL01	191535	13.79				
VN0522WBS01	181924	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.