

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2079
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2078-04
Client ID:	WC-A4-04-CMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Ammonia	mg/L	75-125	1.40		0.42		1	1	98		05/22/2025
ASTM COD	mg/L	75-125	82.5		39.7		50.0	1	86		05/22/2025
ASTM Oil and Grease	mg/L	78-114	20.2		0.29	U	20.0	1	101		05/21/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2079
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2078-04
Client ID:	WC-A4-04-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Ammonia	mg/L	75-125	1.50		0.42		1	1	108		05/22/2025
ASTM COD	mg/L	75-125	80.5		39.7		50.0	1	82		05/22/2025
ASTM Oil and Grease	mg/L	78-114	20.3		0.29	U	20.0	1	102		05/21/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2079
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2079-10
Client ID:	WC-A2-03-CMS	Percent Solids for Spike Sample:	79.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	369		194		125	1	140	*	05/23/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2079
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2079-10
Client ID:	WC-A2-03-CMSD	Percent Solids for Spike Sample:	79.9

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	412		194		125	1	175	*	05/23/2025