

SOP ID : M1311-TCLP-15
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : N/A
Tumbler ID : T-1 / T-2
TCLP Filter ID : 115525

Start Prep Date : 05/20/2025 Time : 17:00
End Prep Date : 05/21/2025 Time : 09:35
Combination Ratio : 20
ZHE Cleaning Batch : ⁷⁸N/A
Initial Room Temperature: 23 °C
Final Room Temperature: 22 °C
TCLP Technician Signature : *JP*
Supervisor By : *12*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 /T-2 checked,30 rpm. q2095-08 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/21/25 12:40	<i>JP</i> / <i>Lab Room</i>	<i>SPG.</i> / <i>RS. / EAT</i>
	Preparation Group	Analysis Group

1 net dry