



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT SDG No.: Q2079
Contract: ENTA05 Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.7	20.0	98	65 - 135	P	05/22/2025	15:30	LB135892
	Barium	109	100	109	65 - 135	P	05/22/2025	15:30	LB135892
	Cadmium	6.20	6.0	103	65 - 135	P	05/22/2025	15:30	LB135892
	Chromium	10.5	10.0	105	65 - 135	P	05/22/2025	15:30	LB135892
	Copper	23.5	20.0	118	65 - 135	P	05/22/2025	15:30	LB135892
	Lead	12.1	12.0	101	65 - 135	P	05/22/2025	15:30	LB135892
	Nickel	41.4	40.0	104	65 - 135	P	05/22/2025	15:30	LB135892
	Selenium	15.4	20.0	77	65 - 135	P	05/22/2025	15:30	LB135892
	Silver	10.1	10.0	101	65 - 135	P	05/22/2025	15:30	LB135892
	Zinc	46.0	40.0	115	65 - 135	P	05/22/2025	15:30	LB135892
CRA	Mercury	0.20	0.2	101	70 - 130	CV	05/23/2025	09:22	LB135895
CRI01	Arsenic	19.9	20.0	100	65 - 135	P	05/28/2025	12:53	LB135935
	Barium	86.9	100	87	65 - 135	P	05/28/2025	12:53	LB135935
	Cadmium	5.78	6.0	96	65 - 135	P	05/28/2025	12:53	LB135935
	Chromium	10.2	10.0	102	65 - 135	P	05/28/2025	12:53	LB135935
	Copper	20.3	20.0	101	65 - 135	P	05/28/2025	12:53	LB135935
	Lead	12.8	12.0	107	65 - 135	P	05/28/2025	12:53	LB135935
	Nickel	38.0	40.0	95	65 - 135	P	05/28/2025	12:53	LB135935
	Selenium	18.1	20.0	90	65 - 135	P	05/28/2025	12:53	LB135935
	Silver	10.3	10.0	103	65 - 135	P	05/28/2025	12:53	LB135935
	Zinc	43.8	40.0	109	65 - 135	P	05/28/2025	12:53	LB135935