

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2079
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2079 **SAS No.:** Q2079

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168113BS							
Arsenic	ug/L	4000	3770		94	80 - 120	P
Barium	ug/L	1000	974		97	80 - 120	P
Cadmium	ug/L	1000	955		96	80 - 120	P
Chromium	ug/L	2000	1990		100	80 - 120	P
Copper	ug/L	1500	1490		99	80 - 120	P
Lead	ug/L	5000	4670		93	80 - 120	P
Nickel	ug/L	2500	2400		96	80 - 120	P
Selenium	ug/L	10000	9550		96	80 - 120	P
Silver	ug/L	380	358		94	80 - 120	P
Zinc	ug/L	1000	1000		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2079
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2079 **SAS No.:** Q2079

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168135BS Mercury	ug/L	4.0	4.04		101	80 - 120	CV