

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WCS-TP2L

Lab Name: Chemtech Consulting Group

Contract: ENTA05

Lab Code: CHEM **Lb No.:** lb135935 **Lab Sample ID :** Q2095-08L **SDG No.:** Q2079

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	1340		1320	J	1		P
Cadmium	20.7	J	150	U	100.0		P
Chromium	16.5	J	250	U	100.0		P
Copper	57.2	J	500	U	100.0		P
Lead	60.0	U	300	U			P
Mercury	2.00	U	10.0	U			CV
Nickel	51.4	J	1000	U	100.0		P
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P
Zinc	298		1000	U	100.0		P