

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

Continuing Calib Date: 05/22/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 03:27 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.65	5.66	5.56	5.76	0.01
Aroclor-1242-2 (2)	5.67	5.68	5.58	5.78	0.01
Aroclor-1242-3 (3)	5.74	5.74	5.64	5.84	0.00
Aroclor-1242-4 (4)	5.83	5.84	5.74	5.94	0.01
Aroclor-1242-5 (5)	6.56	6.57	6.47	6.67	0.01
Tetrachloro-m-xylene	4.50	4.51	4.41	4.61	0.01
Decachlorobiphenyl	10.21	10.22	10.12	10.32	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

Continuing Calib Date: 05/22/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 03:27 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1242-2	(2)	4.90	4.90	4.80	5.00	0.00
Aroclor-1242-3	(3)	5.07	5.08	4.98	5.18	0.01
Aroclor-1242-4	(4)	5.16	5.16	5.06	5.26	0.00
Aroclor-1242-5	(5)	5.68	5.69	5.59	5.79	0.01
Tetrachloro-m-xylene		3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl		8.82	8.83	8.73	8.93	0.01