

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072269.D Time Analyzed: 22:17

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.653  | 5.552     | 5.752  | 474.380            | 500.000           | -5.1 |
| Aroclor-1016-2       | 5.674  | 5.574     | 5.774  | 504.990            | 500.000           | 1.0  |
| Aroclor-1016-3       | 5.737  | 5.636     | 5.836  | 490.580            | 500.000           | -1.9 |
| Aroclor-1016-4       | 5.834  | 5.734     | 5.934  | 501.930            | 500.000           | 0.4  |
| Aroclor-1016-5       | 6.127  | 6.027     | 6.227  | 506.390            | 500.000           | 1.3  |
| Aroclor-1260-1       | 7.245  | 7.147     | 7.347  | 498.780            | 500.000           | -0.2 |
| Aroclor-1260-2       | 7.499  | 7.401     | 7.601  | 477.400            | 500.000           | -4.5 |
| Aroclor-1260-3       | 7.857  | 7.760     | 7.960  | 486.380            | 500.000           | -2.7 |
| Aroclor-1260-4       | 8.082  | 7.985     | 8.185  | 483.510            | 500.000           | -3.3 |
| Aroclor-1260-5       | 8.401  | 8.303     | 8.503  | 498.490            | 500.000           | -0.3 |
| Decachlorobiphenyl   | 10.207 | 10.113    | 10.313 | 50.420             | 50.000            | 0.8  |
| Tetrachloro-m-xylene | 4.500  | 4.396     | 4.596  | 52.020             | 50.000            | 4.0  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP072269.D Time Analyzed: 22:17

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.878 | 4.777                | 4.977 | 497.510            | 500.000           | -0.5 |
| Aroclor-1016-2       | 4.896 | 4.795                | 4.995 | 506.510            | 500.000           | 1.3  |
| Aroclor-1016-3       | 5.073 | 4.973                | 5.173 | 523.110            | 500.000           | 4.6  |
| Aroclor-1016-4       | 5.115 | 5.015                | 5.215 | 521.750            | 500.000           | 4.4  |
| Aroclor-1016-5       | 5.329 | 5.229                | 5.429 | 503.860            | 500.000           | 0.8  |
| Aroclor-1260-1       | 6.362 | 6.264                | 6.464 | 507.640            | 500.000           | 1.5  |
| Aroclor-1260-2       | 6.551 | 6.453                | 6.653 | 486.840            | 500.000           | -2.6 |
| Aroclor-1260-3       | 6.703 | 6.606                | 6.806 | 535.950            | 500.000           | 7.2  |
| Aroclor-1260-4       | 7.174 | 7.077                | 7.277 | 525.860            | 500.000           | 5.2  |
| Aroclor-1260-5       | 7.416 | 7.319                | 7.519 | 538.710            | 500.000           | 7.7  |
| Decachlorobiphenyl   | 8.819 | 8.723                | 8.923 | 53.420             | 50.000            | 6.8  |
| Tetrachloro-m-xylene | 3.795 | 3.691                | 3.891 | 52.290             | 50.000            | 4.6  |