

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030337.D Time Analyzed: 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	8.845	8.750	8.950	735.060	712.500	3.2
2,4-D	8.005	7.910	8.110	719.620	705.000	2.1
2,4-DCAA	6.929	6.834	7.034	763.420	750.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030337.D Time Analyzed: 09:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.512	9.423	9.623	727.860	712.500	2.2
2,4-D	8.635	8.544	8.744	711.970	705.000	1.0
2,4-DCAA	7.447	7.354	7.554	738.530	750.000	-1.5