

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030359.D Time Analyzed: 20:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.844	8.750	8.950	719.260	712.500	0.9
2,4-D	8.004	7.910	8.110	706.710	705.000	0.2
2,4-DCAA	6.929	6.834	7.034	749.220	750.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030359.D Time Analyzed: 20:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.511	9.423	9.623	717.350	712.500	0.7
2,4-D	8.634	8.544	8.744	701.130	705.000	-0.5
2,4-DCAA	7.446	7.354	7.554	730.330	750.000	-2.6