

Surrogate Summary

SDG No.: Q2079

Client: ENTACT

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS030124.D	PIBLK-PS030124.D	2,4-DCAA	1	500	419	84		70 (39)	130 (175)
		2,4-DCAA	2	500	418	84		70 (39)	130 (175)
I.BLK-PS030336.D	PIBLK-PS030336.D	2,4-DCAA	1	500	507	101		70 (39)	130 (175)
		2,4-DCAA	2	500	491	98		70 (39)	130 (175)
Q2080-06MS	PL-HRH-COMP-01MS	2,4-DCAA	1	500	491	98		70 (39)	130 (175)
		2,4-DCAA	2	500	421	84		70 (39)	130 (175)
Q2080-06MSD	PL-HRH-COMP-01MSD	2,4-DCAA	1	500	496	99		70 (39)	130 (175)
		2,4-DCAA	2	500	428	86		70 (39)	130 (175)
PB168120BL	PB168120BL	2,4-DCAA	1	500	470	94		70 (39)	130 (175)
		2,4-DCAA	2	500	455	91		70 (39)	130 (175)
PB168120BS	PB168120BS	2,4-DCAA	1	500	550	110		70 (39)	130 (175)
		2,4-DCAA	2	500	529	106		70 (39)	130 (175)
PB168092TB	PB168092TB	2,4-DCAA	1	500	715	143	*	70 (39)	130 (175)
		2,4-DCAA	2	500	482	96		70 (39)	130 (175)
I.BLK-PS030347.D	PIBLK-PS030347.D	2,4-DCAA	1	500	518	104		70 (39)	130 (175)
		2,4-DCAA	2	500	510	102		70 (39)	130 (175)
Q2079-03	WC-A2-01-C	2,4-DCAA	1	500	617	123		70 (39)	130 (175)
		2,4-DCAA	2	500	443	89		70 (39)	130 (175)
Q2079-07	WC-A2-02-C	2,4-DCAA	1	500	808	162	*	70 (39)	130 (175)
		2,4-DCAA	2	500	594	119		70 (39)	130 (175)
Q2079-11	WC-A2-03-C	2,4-DCAA	1	500	608	122		70 (39)	130 (175)
		2,4-DCAA	2	500	455	91		70 (39)	130 (175)
I.BLK-PS030358.D	PIBLK-PS030358.D	2,4-DCAA	1	500	513	103		70 (39)	130 (175)
		2,4-DCAA	2	500	501	100		70 (39)	130 (175)