

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2079

**Analytical Method:** 8151A

**Client:** ENTACT

**Datafile :** PS030344.D

| Lab Sample ID            | Parameter        | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|--------------------------|------------------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|                          |                  |       |        |       |     |     |      | Qual | Low     | High      |     |
| PB168120BS<br>(Column 1) | 2,4-D            | 5     | 5.20   | ug/L  | 104 |     |      |      | 70 (83) | 130 (130) |     |
|                          | 2,4,5-TP(Silvex) | 5     | 5.20   | ug/L  | 104 |     |      |      | 70 (78) | 130 (127) |     |
| PB168120BS<br>(Column 2) | 2,4-D            | 5     | 5.10   | ug/L  | 102 |     |      |      | 70 (83) | 130 (130) |     |
|                          | 2,4,5-TP(Silvex) | 5     | 5.10   | ug/L  | 102 |     |      |      | 70 (78) | 130 (127) |     |