

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q2079 **SAS No.:** Q2079 **SDG NO.:** Q2079

Instrument ID: ECD_L **Date(s) Analyzed:** 05/21/2025 05/21/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.74	4.64	4.84	175163000
		2	5.27	5.17	5.37	186705000
		3	5.97	5.87	6.07	729383000
		4	6.06	5.96	6.16	887825000
		5	6.90	6.80	7.00	133478000
Toxaphene	500	1	5.88	5.78	5.98	11156600
		2	6.27	6.17	6.37	26557800
		3	7.09	6.99	7.19	93865100
		4	7.18	7.08	7.28	65544600
		5	7.96	7.86	8.06	45659000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q2079 **SAS No.:** Q2079 **SDG NO.:** Q2079

Instrument ID: ECD_L **Date(s) Analyzed:** 05/21/2025 05/21/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.91	3.81	4.01	190212000
		2	4.49	4.39	4.59	216767000
		3	5.12	5.02	5.22	652504000
		4	5.19	5.09	5.29	564939000
		5	6.08	5.98	6.18	249456000
Toxaphene	500	1	5.15	5.05	5.25	30765000
		2	5.83	5.73	5.93	34678400
		3	6.11	6.01	6.21	36071200
		4	6.75	6.65	6.85	117492000
		5	7.19	7.09	7.29	80236600