

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095768.D Time Analyzed: 13:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.102	9.003	9.203	50.550	50.000	1.1
Endrin	6.602	6.502	6.702	44.840	50.000	-10.3
gamma-BHC (Lindane)	4.355	4.257	4.457	54.120	50.000	8.2
Heptachlor	4.953	4.855	5.055	50.760	50.000	1.5
Heptachlor epoxide	5.717	5.618	5.818	54.480	50.000	9.0
Methoxychlor	7.522	7.423	7.623	42.480	50.000	-15.0
Tetrachloro-m-xylene	3.572	3.473	3.673	54.260	50.000	8.5

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/22/2025

Lab Sample No.: PSTDCCC050 Data File : PL095768.D Time Analyzed: 13:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.964	8.164	54.400	50.000	8.8
Endrin	5.781	5.683	5.883	43.710	50.000	-12.6
gamma-BHC (Lindane)	3.729	3.630	3.830	53.430	50.000	6.9
Heptachlor	4.082	3.983	4.183	49.240	50.000	-1.5
Heptachlor epoxide	4.869	4.771	4.971	52.730	50.000	5.5
Methoxychlor	6.749	6.651	6.851	41.100	50.000	-17.8
Tetrachloro-m-xylene	2.885	2.786	2.986	52.800	50.000	5.6