

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2079

Analytical Method: 8081B

Client: ENTACT

Datafile : PL095762.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168121BS (Column 1)	gamma-BHC (Lindane)	0.5	0.57	ug/L	114				40 (82)	140 (129)	
	Heptachlor	0.5	0.54	ug/L	107				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.58	ug/L	117				40 (81)	140 (124)	
	Endrin	0.5	0.47	ug/L	94				40 (81)	140 (128)	
	Methoxychlor	0.5	0.44	ug/L	87				40 (78)	140 (108)	
PB168121BS (Column 2)	gamma-BHC (Lindane)	0.5	0.55	ug/L	110				40 (82)	140 (129)	
	Heptachlor	0.5	0.51	ug/L	102				40 (79)	140 (127)	
	Heptachlor epoxide	0.5	0.56	ug/L	111				40 (81)	140 (124)	
	Endrin	0.5	0.45	ug/L	91				40 (81)	140 (128)	
	Methoxychlor	0.5	0.40	ug/L	81				40 (78)	140 (108)	