

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2079

Client: ENTACT

Analytical Method: 8270E DataFile: BP024790.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168131BS	Pyridine	50	33.8	ug/L	68				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	42.1	ug/L	84				70 (76)	130 (103)	
	2-Methylphenol	50	42.6	ug/L	85				70 (69)	130 (109)	
	3+4-Methylphenols	50	41.9	ug/L	84				20 (67)	160 (106)	
	Hexachloroethane	50	40.2	ug/L	80				20 (76)	160 (118)	
	Nitrobenzene	50	39.7	ug/L	79				70 (58)	130 (106)	
	Hexachlorobutadiene	50	46.4	ug/L	93				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	48.0	ug/L	96				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	47.8	ug/L	96				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	47.7	ug/L	95				70 (60)	130 (115)	
	Hexachlorobenzene	50	47.0	ug/L	94				70 (73)	130 (106)	
	Pentachlorophenol	100	110	ug/L	110				20 (47)	160 (114)	