

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM

SAS No.: Q2079

SDG NO.: Q2079

Lab File ID: BP024787.D

DFTPP Injection Date: 05/23/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.4
68	Less than 2.0% of mass 69	0.1 (0.7) 1
69	Mass 69 relative abundance	20.9
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	30.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	4.7
275	10.0 - 60.0% of mass 198	23.1
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	15.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.8 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024788.D	05/23/2025	11:40
PB168131BL	PB168131BL	BP024789.D	05/23/2025	12:20
PB168131BS	PB168131BS	BP024790.D	05/23/2025	13:01