

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG No.: Q2079

Instrument ID: MSVOA_N Calibration Date/Time: 05/22/2025 08:21

Lab File ID: VN086750.D Init. Calib. Date(s): 05/16/2025 05/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 16:31 18:32

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.733	0.639		-12.82	20
1,1-Dichloroethene	0.558	0.523		-6.27	20
2-Butanone	0.399	0.327		-18.05	20
Carbon Tetrachloride	0.433	0.429		-0.92	20
Chloroform	1.119	0.975		-12.87	20
Benzene	1.433	1.359		-5.16	20
1,2-Dichloroethane	0.446	0.396		-11.21	20
Trichloroethene	0.346	0.360		4.05	20
Tetrachloroethene	0.383	0.420		9.66	20
Chlorobenzene	1.096	1.093	0.3	-0.27	20
1,2-Dichloroethane-d4	0.691	0.561		-18.81	20
Dibromofluoromethane	0.298	0.288		-3.36	20
Toluene-d8	1.234	1.187		-3.81	20
4-Bromofluorobenzene	0.436	0.423		-2.98	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.