

Surrogate Summary

SDG No.: Q2079

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2079-01	WC-A2-01-G	1,2-Dichloroethane-d4	50	43.6	87	70 (74)	130 (125)
		Dibromofluoromethane	50	33.9	68 *	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.7	107	70 (77)	130 (121)
Q2079-01RE	WC-A2-01-GRE	1,2-Dichloroethane-d4	50	44.1	88	70 (74)	130 (125)
		Dibromofluoromethane	50	33.7	67 *	70 (75)	130 (124)
		Toluene-d8	50	51.8	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.2	110	70 (77)	130 (121)
Q2079-05	WC-A2-02-G	1,2-Dichloroethane-d4	50	43.6	87	70 (74)	130 (125)
		Dibromofluoromethane	50	36.3	73	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.0	112	70 (77)	130 (121)
Q2079-05RE	WC-A2-02-GRE	1,2-Dichloroethane-d4	50	44.9	90	70 (74)	130 (125)
		Dibromofluoromethane	50	36.8	74	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.4	111	70 (77)	130 (121)
Q2079-09	WC-A2-03-G	1,2-Dichloroethane-d4	50	44.5	89	70 (74)	130 (125)
		Dibromofluoromethane	50	39.1	78	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.3	109	70 (77)	130 (121)
VN0521WBL01	VN0521WBL01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	53.5	107	70 (75)	130 (124)
		Toluene-d8	50	52.6	105	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108	70 (77)	130 (121)
VN0521WBS01	VN0521WBS01	1,2-Dichloroethane-d4	50	53.9	108	70 (74)	130 (125)
		Dibromofluoromethane	50	53.2	106	70 (75)	130 (124)
		Toluene-d8	50	53.1	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.8	108	70 (77)	130 (121)
VN0521WBSD0	VN0521WBSD01	1,2-Dichloroethane-d4	50	50.0	100	70 (74)	130 (125)
		Dibromofluoromethane	50	51.2	102	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)
VN0522WBL01	VN0522WBL01	1,2-Dichloroethane-d4	50	44.2	88	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VN0522WBS01	VN0522WBS01	1,2-Dichloroethane-d4	50	42.1	84	70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100	70 (75)	130 (124)
		Toluene-d8	50	49.3	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97	70 (77)	130 (121)

