

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079
 Lab File ID: VN086722.D Date Analyzed: 05/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:10
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	182299	8.22	331408	9.10	299657	11.87
UPPER LIMIT	364598	8.724	662816	9.6	599314	12.365
LOWER LIMIT	91149.5	7.724	165704	8.6	149829	11.365
EPA SAMPLE NO.						
WC-A2-01-G	247245	8.22	450917	9.10	434993	11.86
WC-A2-02-G	257060	8.22	475048	9.10	461977	11.86
WC-A2-03-G	258359	8.22	480749	9.10	461451	11.86
VN0521WBL01	174294	8.22	341228	9.10	334894	11.87
VN0521WBS01	167977	8.22	308282	9.10	282384	11.87
VN0521WBSD01	190593	8.22	345539	9.10	308855	11.86

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.