

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2079 SAS No.: Q2079 SDG NO.: Q2079
 Lab File ID: VN086750.D Date Analyzed: 05/22/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	268645	8.22	454328	9.09	397309	11.87
UPPER LIMIT	537290	8.724	908656	9.594	794618	12.365
LOWER LIMIT	134323	7.724	227164	8.594	198655	11.365
EPA SAMPLE NO.						
WC-A2-01-GRE	240765	8.22	441543	9.09	433303	11.87
WC-A2-02-GRE	240014	8.22	448142	9.10	429564	11.87
VN0522WBL01	246936	8.22	454612	9.10	435552	11.87
VN0522WBS01	281968	8.22	476790	9.10	414976	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.