



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**ATC Group Services LLC**

**Project Name: PS 152 Brooklyn**

**Project # N/A**

**Order ID # Q2081**

**Test Name: PCB Group1**

### **A. Number of Samples and Date of Receipt:**

1 Solid sample was received on 05/19/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: PCB Group1. This data package contains results for PCB Group1.

### **C. Analytical Techniques:**

The analyses were performed on instrument GCECD\_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCB Group1s was based on method 8082A and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

Samples CAULK was diluted due to Caulk matrix.

### **E. Additional Comments:**

The temperature of the samples at the time of receipt was 11.2°C. No MSMSD performed as samples are caulk matrix.

Samples were received on 05/19/2025, 14:155 and composited in the Lab on 05/19/2025, 15:56

### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_