

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

Continuing Calib Date: 05/20/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 22:08 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.77       | 4.77      | 4.67      | 4.87 | 0.00       |
| Aroclor-1242-2       | (2) | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Aroclor-1242-3       | (3) | 4.85       | 4.85      | 4.75      | 4.95 | 0.00       |
| Aroclor-1242-4       | (4) | 4.97       | 4.97      | 4.87      | 5.07 | 0.00       |
| Aroclor-1242-5       | (5) | 5.62       | 5.62      | 5.52      | 5.72 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.72       | 8.73      | 8.63      | 8.83 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

Continuing Calib Date: 05/20/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 22:08 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.76       | 4.76      | 4.66      | 4.86 | 0.00       |
| Aroclor-1242-2       | (2) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-3       | (3) | 4.95       | 4.95      | 4.85      | 5.05 | 0.00       |
| Aroclor-1242-4       | (4) | 5.04       | 5.04      | 4.94      | 5.14 | 0.01       |
| Aroclor-1242-5       | (5) | 5.56       | 5.56      | 5.46      | 5.66 | 0.01       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.67       | 8.67      | 8.57      | 8.77 | 0.00       |