

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PO111211.D Time Analyzed: 10:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.673	4.873	504.450	500.000	0.9
Aroclor-1016-2	4.795	4.692	4.892	507.350	500.000	1.5
Aroclor-1016-3	4.852	4.749	4.949	516.940	500.000	3.4
Aroclor-1016-4	4.972	4.869	5.069	507.460	500.000	1.5
Aroclor-1016-5	5.230	5.126	5.326	514.510	500.000	2.9
Aroclor-1260-1	6.269	6.166	6.366	512.080	500.000	2.4
Aroclor-1260-2	6.459	6.355	6.555	509.170	500.000	1.8
Aroclor-1260-3	6.826	6.722	6.922	505.340	500.000	1.1
Aroclor-1260-4	7.086	6.983	7.183	503.850	500.000	0.8
Aroclor-1260-5	7.329	7.225	7.425	504.320	500.000	0.9
Decachlorobiphenyl	8.728	8.624	8.824	49.320	50.000	-1.4
Tetrachloro-m-xylene	3.686	3.582	3.782	50.400	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PO111211.D Time Analyzed: 10:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.658	4.858	456.250	500.000	-8.8
Aroclor-1016-2	4.776	4.677	4.877	501.670	500.000	0.3
Aroclor-1016-3	4.952	4.852	5.052	490.490	500.000	-1.9
Aroclor-1016-4	4.994	4.894	5.094	485.960	500.000	-2.8
Aroclor-1016-5	5.207	5.107	5.307	487.900	500.000	-2.4
Aroclor-1260-1	6.237	6.138	6.338	494.400	500.000	-1.1
Aroclor-1260-2	6.425	6.326	6.526	489.530	500.000	-2.1
Aroclor-1260-3	6.578	6.479	6.679	459.900	500.000	-8.0
Aroclor-1260-4	7.049	6.949	7.149	491.950	500.000	-1.6
Aroclor-1260-5	7.290	7.191	7.391	487.210	500.000	-2.6
Decachlorobiphenyl	8.675	8.574	8.774	48.440	50.000	-3.1
Tetrachloro-m-xylene	3.678	3.579	3.779	49.320	50.000	-1.4