

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PO111226.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.771	4.673	4.873	501.380	500.000	0.3
Aroclor-1016-2	4.790	4.692	4.892	505.310	500.000	1.1
Aroclor-1016-3	4.847	4.749	4.949	518.240	500.000	3.6
Aroclor-1016-4	4.967	4.869	5.069	507.270	500.000	1.5
Aroclor-1016-5	5.225	5.126	5.326	507.680	500.000	1.5
Aroclor-1260-1	6.264	6.166	6.366	497.960	500.000	-0.4
Aroclor-1260-2	6.453	6.355	6.555	501.710	500.000	0.3
Aroclor-1260-3	6.820	6.722	6.922	485.860	500.000	-2.8
Aroclor-1260-4	7.080	6.983	7.183	470.870	500.000	-5.8
Aroclor-1260-5	7.322	7.225	7.425	462.400	500.000	-7.5
Decachlorobiphenyl	8.721	8.624	8.824	45.530	50.000	-8.9
Tetrachloro-m-xylene	3.681	3.582	3.782	50.530	50.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q2081 SAS No.: Q2081 SDG NO.: Q2081

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PO111226.D Time Analyzed: 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.658	4.858	480.240	500.000	-4.0
Aroclor-1016-2	4.776	4.677	4.877	482.760	500.000	-3.4
Aroclor-1016-3	4.951	4.852	5.052	487.320	500.000	-2.5
Aroclor-1016-4	4.993	4.894	5.094	483.190	500.000	-3.4
Aroclor-1016-5	5.206	5.107	5.307	486.570	500.000	-2.7
Aroclor-1260-1	6.236	6.138	6.338	470.860	500.000	-5.8
Aroclor-1260-2	6.424	6.326	6.526	462.740	500.000	-7.5
Aroclor-1260-3	6.577	6.479	6.679	431.340	500.000	-13.7
Aroclor-1260-4	7.047	6.949	7.149	452.800	500.000	-9.4
Aroclor-1260-5	7.288	7.191	7.391	440.990	500.000	-11.8
Decachlorobiphenyl	8.672	8.574	8.774	44.770	50.000	-10.5
Tetrachloro-m-xylene	3.678	3.579	3.779	49.760	50.000	-0.5