

Surrogate Summary

SDG No.: Q2081

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111057.D	PIBLK-PO111057.D	Tetrachloro-m-xylene	1	20	16.4	82		60	140
		Decachlorobiphenyl	1	20	17.5	87		60	140
		Tetrachloro-m-xylene	2	20	15.4	77		60	140
		Decachlorobiphenyl	2	20	17.9	90		60	140
I.BLK-PO111215.D	PIBLK-PO111215.D	Tetrachloro-m-xylene	1	20	18.9	95		16	158
		Decachlorobiphenyl	1	20	19.3	97		10	173
		Tetrachloro-m-xylene	2	20	17.1	86		16	158
		Decachlorobiphenyl	2	20	19.4	97		10	173
PB168077BL	PB168077BL	Tetrachloro-m-xylene	1	20	19.9	100		32	144
		Decachlorobiphenyl	1	20	20.2	101		32	175
		Tetrachloro-m-xylene	2	20	17.7	89		32	144
		Decachlorobiphenyl	2	20	19.9	99		32	175
PB168077BS	PB168077BS	Tetrachloro-m-xylene	1	20	19.1	96		32	144
		Decachlorobiphenyl	1	20	21.2	106		32	175
		Tetrachloro-m-xylene	2	20	17.6	88		32	144
		Decachlorobiphenyl	2	20	20.9	105		32	175
Q2081-01	CAULK	Tetrachloro-m-xylene	1	20	16.3	81		32	144
		Decachlorobiphenyl	1	20	8.80	44		32	175
		Tetrachloro-m-xylene	2	20	11.2	56		32	144
		Decachlorobiphenyl	2	20	8.90	45		32	175
I.BLK-PO111230.D	PIBLK-PO111230.D	Tetrachloro-m-xylene	1	20	19.0	95		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	17.3	86		60	140
		Decachlorobiphenyl	2	20	18.5	92		60	140