

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2081</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ATC Group Services LLC</u>	Datafile :	<u>PO111217.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168077BS (Column 1)	AR1016	166.5	155	ug/kg	93				71	120		
	AR1260	166.5	151	ug/kg	91				65	130		
	Total PCBs	0	292	ug/kg					0	0		
PB168077BS (Column 2)	AR1016	166.5	145	ug/kg	87				71	120		
	AR1260	166.5	147	ug/kg	88				65	130		
	Total PCBs	0	306	ug/kg					0	0		