

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2082 SAS No.: Q2082 SDG NO.: Q2082
Lab File ID: VX046284.D Date Analyzed: 05/21/2025
Instrument ID: MSVOA_X Time Analyzed: 09:55
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	67683	12.018				
UPPER LIMIT	135366	12.518				
LOWER LIMIT	33841.5	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250515	60241	12.02				
BP-VPB-182-GW-260-262	51747	12.02				
BP-VPB-182-GW-280-282	47547	12.02				
BP-VPB-182-GW-300-302	58719	12.02				
BP-VPB-182-GW-320-322	48477	12.02				
BP-VPB-182-GW-360-362	50238	12.02				
BP-VPB-182-GW-390-392	52601	12.02				
BP-VPB-182-GW-400-402	53505	12.02				
BP-VPB-182-GW-400-402MS	57803	12.02				
BP-VPB-182-GW-400-402MSD	59572	12.02				
VPB182-HYD-20250516	54802	12.02				
VX0521WBL01	53099	12.02				
VX0521WBS01	62651	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.